



# MEDICAL TERMINAL UNITS

# **GCE MEDIUNIT-SS TYPE**

## **EDITION 1/2014**

#### **MEDICAL TERMINAL UNITS**

Medical terminal units provide quick and easy connection of hospital ward gas equipment to the hospital source.

GCE complies with ISO 7396 and national installation standards with secure products where every product is fully tested in production. Our Medical gas outlets are in accordance with ISO EN 9170-1 international standards.

### **KEY BENEFITS**

- Wall housing is compatible with all GCE MediUnit standards like DIN, BSI, SS, CZ
- After 10 years it is possible to upgrade the units with a special upgrade pack
- Small size
- All functional components from brass
- Simple installation
- Easy to clean
- Fast connection and disconnection
- Attractive design
- Recessed, Exposed and Bed head installation versions
- Complies with colour coding and description by standard
- In accordance with ISO 9170-1
- Ultra safe product where every piece is fully tested in production



Recessed version



**Exposed version** 




# **MEDICAL TERMINAL UNIT - SS**

# TECHNICAL DATA

Gases:	O <sub>2</sub> ; N <sub>2</sub> O; Air; Air-800; CO <sub>2</sub> ; N <sub>2</sub> ; O <sub>2</sub> /N <sub>2</sub> O; AGSS; Ar; VAC
Dimensions:	Height: 73 mm, Width: 73 mm, Depth: 63 mm
Weight:	0,36 kg
Working pressure:	General: 4 - 5 bar
	AIR-800: 7 - 10 bar
	VAC: (-0,4) - (-0,9) bar
Regulatory status:	Complies with Medical directive 93/42/EEC
	Complies with EN ISO 7396-1 (Central Gas Supply system)
	Complies with EN ISO 9170-1 (Terminal units)
	Complies with SS 8752430 (SS gas specific connections)

### **MEASUREMENTS**



## **SPLIT VIEW**





Installation plug

## **INSTALLATION TOOLS**

MP_00345	QC installation key
MP_00324	Button remover
0732040	Installation plug (10 pcs)

## MOUNTING



- Gas specific connection Probe Terminal unit
- Colour coding marking on terminal unit



- Push probe to working position like on the picture
- In working position Probe is connected to the gas supply
- Pressure drop is comply with EN 9170-1



• Press the button to remove probe

### **GCE Norden AB**

Källvattengatan 9, Box 21044, SE-200 21 Malmö Telephone: +46 (0) 40 38 83 00, Fax: +46 (0) 40 38 83 53 www.gcegroup.com

